

EU-DECLARATION OF CONFORMITY

Manufacturer Wero Swiss Med Kft.
Ipartelep utca 6
4220 Hajdúböszörmény
Hungary

Single Registration Number: HU-MF-000006727

We declare under our sole responsibility that

the medical device **Urgo Last Color**
with REF 510745-47 / 510765-67 and
510771-76
with Basic-UDI-DI (productcode) **76305236A14S9**

Risk Class class 1, unsterile
According to Annex VIII, Rule 1


meets all the provisions of the Regulation (EU) 2017 /745 which apply to it.

Applied harmonised standards, national standards or other normative documents EN ISO 10993-1:2020
EN ISO 15223-1:2017
EN ISO 14971:2019

Conformity assessment procedure: Regulation (EU) 2017/745
according to Annex IV

Place, date Hajdúböszörmény, 22.03.2022

Name and function


Tibor Balla, Head of Quality Management

