

EC Declaration of Conformity

Manufacturer:

Jiangyin Everise Medical Equipment Co.,Ltd
1001 Chengyang Road,Jiangyin City,
Jiangsu,China,

European Representative:

Share Info Gmbh
Heerdter Lohweg 83, 40549 Düsseldorf

We, the manufacturer, herewith declare that the products:

Medical Stretchers
See the Annex I for details

Basic UDI: 697645837WSX-AWY
SRN: CN-MF-000010106
EMDN code: V0899

Classification (MDR, Annex VIII): Class I, Rule 1 according to regulation (EU) 2017/745

Meets the provisions of the Regulation EU 2017/745(MDR).The declaration is valid in connection with the “final inspection report” of the device.

Conformity Assessment Route: Annex II and Annex III, regulation (EU) 2017/745


We herewith declare that the above mentioned products meet the transposition into national law, the provisions of the following EC Council Directives and Standards. All supporting documentations are retained under the premises of the manufacturer. We are exclusively responsible for the Declaration of Conformity.

The above mentioned declaration of conformity is exclusively under the responsibility of

Jiangyin Everise Medical Equipment Co.,Ltd
1001 Chengyang Road,Jiangyin City,Jiangsu,China,

Wuxi City, 2021.05

Place, date



Name and function

Annex I: Model and specification list

Product name	Model	Length (mm)	Width (mm)	Height (mm)
Stretcher cot	WSX-A1	1900	550	840
	WSX-A2	1900	550	830
	WSX-A3	1980	550	870
	WSX-A5	1940	525	810
	WSX-A7	1975	530	885
	WSX-A9	1960	545	≥800
	WSX-A11	2030	560	860
	WSX-A15	≥1900	≥540	≥800
	WSX-A8	1900	570	910
	WSX-A8N	1935	570	915
	WSX-A20	≥1950	≥550	≥890
WSX-A10	1972	542	≥800	
Medical Trolley	WSX-B1-1	1930	550	412
	WSX-B1-2	1930	550	312
Folding Stretcher	WSX-C1	1885	485	190
	WSX-C2	1885	485	30
	WSX-C3	2065	530	185
	WSX-C4	1900	475	200
	WSX-C6	1900	475	170
	WSX-C1B	2050	530	135
	WSX-D1	2050	530	130
	WSX-D2	2090	535	110
	WSX-D1A	2055	515	130
	WSX-D4	2030	530	155
	WSX-D1C	2050	530	135
	WSX-D5	2290	570	145
WSX-D8	2290	570	145	
Scoop Stretcher	WSX-E3	1710	432	85
	WSX-E4	1700	432	85
	WSX-E5	1640	445	65
	WSX-E6	1705	430	70
	WSX-E8	1640	435	70
	WSX-E8S	1723	436	73
Stair Chair	WSX-G1	710	505	910
	WSX-G2	780	505	980
Patient Transfer	WSX-S	1815	445	55
Patient Transfer Pad	WSX-L1	1980	970	/
	WSX-L2	1916	515	/
Basket Stretcher	WSX-R1	2232	622	170